

KETERANGAN PEMBELIAN TEMPATAN - JKP KEMENTAH

RUJUKAN SEBUTHARGA : MINDEF/DFA/FY2026/NAU/DEN/LP/1004

TARIKH TUTUP SEBUTHARGA : 20 - 8 - 2026 (KHAMIS, 08:00 PAGI)

PETI PENERIMAAN SEBUTHARGA : BOX (3)

ALAMAT BARU: BANGUNAN ARKIB, ARAS BAWAH, KEMENTERIAN PERTAHANAN,
BOLKIAH GARISON

NAMA SYARIKAT : _____

CAP SYARIKAT DAN TANDATANGAN : _____

TEL PEMBEKAL : _____ FAX PEMBEKAL : _____

E-MAIL PEMBEKAL : _____

VENDOR ID PEMBEKAL : _____

BIL	KETERANGAN BARANG	KTI	HARGA SATU	HARGA SEMUA	TEMPOH SAH HARGA (MINIMUM 180 HARI)	TEMPOH PENGHANTARAN	BUATAN NEGERI	LAIN-LAIN HAL
1	<u>PEMBELIAN PERALATAN DAN KELEGGAPAN PERGIGIAN (DENTAL EQUIPMENT)</u> DENTAL INTRAORAL SCANNER EQUIPMENT (Sila rujuk spesifikasi di sertakan)	01 UNIT						
PERHATIAN: Perkara-perkara yang perlu (mandatori) di hadapkan sebelum tarikh tutup sebutharga : ☒ Katalog & Spesifikasi Teknikal mesti di sertakan untuk penilaian. ☐ Contoh Jika di dapati gagal, sebutharga tersebut tidak akan di pertimbangkan. PERINGATAN: Tidak dibenarkan menggunakan katalog Jabatanarah ini.								

JUMLAH SEMUA HARGA YANG DI TAWARKAN - ANGKA B\$ _____ DAN PERKATAAN: _____

Dental Intraoral Scanner Equipment	
Demand Reference	DSE/06/26/07
Description	Digital Intraoral Scanner System
Quantity Required	1 unit

Ser	Technical Data & Dimensional		Offer
	Item	Specification	
(a)	(b)	(c)	(d)
Essential			
1	General	An intraoral scanner is a digital device used in dentistry to capture highly accurate 3D images of a patient's teeth, gums, and oral structures without the need for traditional impression materials.	
2	Technical data	Trios 3 Wired 3 shape: Scanner Weight 340 g (Wired) Dimensions (W x H x L) 4.8 x 4.0 x 27.3 cm Tip Dimensions (W x H x L) 2.05 x 2.0 x 8.6 cm Capture Rate Up to 1,875 images per second Light Source LED Connectivity/Power Fully PC-powered Trios Move+ Screen 15.6" Full HD touchscreen Processor Intel Core i5-12600T Operating System Windows 11 IoT Storage 1TB SSD GPU Nvidia T1000 8GB GDDR6 RAM 32GB DDR4 Wheels 4 wheels w. integrated brakes Weight 13 kg	
3	Features	Advantages Improved patient comfort Faster impression-taking process High accuracy and precision Immediate digital transfer to laboratories Reduced material and storage costs Better patient engagement and case presentation Features Proven Clinical Accuracy Real Color & Shade Measurement AI Scan 2.0 Open System Output Patients Engagement Apps Invisalign Connection	



		- Manufacturers authorized the supplier of the equipment - Are authorized by the manufacturer to install; test and commission their equipment	
6	Warranty	It is essential that the manufacturer provided a minimum of 1 year warranty support (parts and labor) from the date of acceptance into service	
7	Training	It is essential that the Supplier provide end-user training on clinical equipment and use of software on patients.	
Desirable			
8	Preferred Model/ Brand & country Origin of the manufacturer	Trios 3 Wired 3 shape with Trios Move+ or Equivalent	

*Important Notice

This specification annex form must complete and attached with the quotation tenders provided.


Company / Supplier Chop & Signature

.....



Dental Services RBAF Demand Catalogue Reference

Appendix 8-1



Dental Intraoral Scanner Equipment	
Dental Services RBAF Demand Reference	DSE/06/26/07
Descriptions	Digital Intraoral Scanner System
Quantity Required	1 Unit
Preferred Model/Brand & Country Origin of the Manufacturer	Preferred Brand: Trios 3 Wired 3 shape with Move+ or Equivalent
Submitting Catalogue/Brochures	The Supplier must submit the Catalogue/brochures completely for the offered item.

