

KETERANGAN PEMBELIAN TEMPATAN - JKP KEMANTAN

RUJUKAN SEBUTHARGA : MINDEF/DFA/FY2026/GCAU/MED/LP/2071
TARIKH TUTUP SEBUTHARGA : 17 / 9 / 2026 (KHAMIS, 8.00 PAGI)
PETI PENERIMAAN SEBUTHARGA : BOX (3)
ALAMAT BARU: : ARAS BAWAH, BANGUNAN ARKIB
KEMENTERIAN PERTAHANAN, BOLKIAH GARISSON

NAMA SYARIKAT : _____
COP SYARIKAT & TANDA TANGAN : _____
TEL PEMBEKAL : _____ FAX PEMBEKAL : _____
EMAIL PEMBEKAL : _____
TAFIS VENDOR ID: _____

BIL	KETERANGAN BARANG	KTI	HARGA SATU	HARGA SEMUA	TEMPOH SAH HARGA (MINIMUM 180 HARI)	TEMPOH PENGHANTARAN	BUATAN NEGERI	TEMPOH HAYAT PENYIMPANAN (SHELF LIFE)	LAIN-LAIN HAL
	PEMBELIAN BARANGAN PERUBATAN BAGI MRS								
1	I.V Surrlo Catherer (18G x 1 ¼") Packaging: Box of 50 Pcs	100 Boxes							
2	I.V Surrlo Catherer (20G x 1") Packaging: Box of 50 Pcs	100 Boxes							
3	I.V Surrlo Catherer (22G x 1") Packaging: Box of 50 Pcs	100 Boxes							
4	I.V Surrlo Catherer (24G x 3/4") Packaging: Box of 50 Pcs	100 Boxes							
A	<i>As per specifications included.</i>								
B	PERINGATAN: Tidak di benarkan menggunakan katalog Jabatanarah ini. PERHATIAN: Perkara-perkara yang perlu (mandatori) dihadapkan sebelum tarikh tutup sebutbarga: ☐ / Katalog ☐ Contoh Jika didapati gagal, sebutbarga tersebut tidak akan dipertimbangkan.								

10
378

JUMLAH SEMUA HARGA YANG DITAWARKAN - ANGKA B\$ _____ DAN PERKATAAN : _____

SPESIFIKASI ALAT PERUBATAN

BIL	ITEM	SPESIFIKASI	JUMLAH
(a)	(b)	(c)	(d)
1	I.V. Surflo Catheter, 18G x 1¼" 	Gauge & Length: 18G x 1¼" Preferred Brand: Terumo or Equivalent Color Code: Green Product Code: SR-OX1832CA Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging Flow Rate: 100 mL / min Sterility: Sterile, disposable, and tamper-evident packaging Radiopaque: X-ray detectable	100 Boxes (box of 50 pcs)
2	I.V. Surflo Catheter, 20G x 1" 	Gauge & Length: 20G x 1" Preferred Brand: Terumo or Equivalent Color Code: Pink Product Code: SR-OX2025CA Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma • Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging Flow Rate: 100 mL / min Sterility: Sterile, disposable, and tamper-evident packaging • Radiopaque: X-ray detectable	100 Boxes (box of 50 pcs)
3	I.V. Surflo Catheter, 22G x 1" 	Gauge & Length: 22G x 1" Preferred Brand: Terumo or Equivalent Color Code: Blue Product Code: SR-OX2225CA Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma • Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging Flow Rate: 100 mL / min Sterility: Sterile, disposable, and tamper-evident packaging • Radiopaque: X-ray detectable	100 Boxes (box of 50 pcs)

4	I.V. Surflo Catheter, 24G x 3/4" 	Gauge & Length: 24G x 3/4" Preffered Brand: Terumo or Equivalent Color Code: Yellow Product Code: SR-OX2419CA Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma • Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging Flow Rate: 100 mL / min Sterility: Sterile, disposable, and tamper-evident packaging Radiopaque: X-ray detectable	100 Boxes (box of 50 pcs)
---	--	---	--------------------------------------

CERAIAN 'C' KPD:

HQ/MED/Q/052

BTKH: 09 SAFAR 1448

24 JULAI 2026

STOR LOGISTIK PERUBATAN BOLKIAH GARISON
COMPLIANCE AGREEMENT TABLE FOR SUPPLIER

Serial *	Item Name			
		I.V. Surflo Catheter, 18G x 1¼"		
	Catalogue / Specification	As Attached		
(a)	(b)	(c)	(d)	(e)
	DESCRIPTION & USER REQUIREMENTS	Comply (√)	Not Comply (√)	Remark
1	Distributor Authorization Letter by the sole agency agreement from the agent			
2	Must provide manual instruction from the agent supply			
3	Clear Catalogue / Specification attached			
4	Certificate Internal Safety and Quality Standard: The supplier should specific the Product is approved by the international safety standards - TUV,FDA, CE MARK, TGA, CSA, SMG, ADA, CGSB, NF, UKAS, MS and Japanese Standard			
5	Attached SAMPLE from vendor			
	SPECIFICATION:			
6	Gauge & Length: 18G × 1¼"			
7	Color Code: Green			
8	Preffered Brand: Terumo or Equivalent			
9	Product Code: SR-OX1832CA			
10	Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips			
11	Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma			
12	Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging			
13	Shelf life: Minimum 3–5 years from the date of manufacture.			
14	Quantity: 100 Boxes (box of 50 pcs)			
15	Delivery Period/Time			
16	Country Origin			
17	Quality Assurance			
Note:				
a. The product will be rejected if it does not match the specifications.				

STOR LOGISTIK PERUBATAN BOLKIAH GARISON
COMPLIANCE AGREEMENT TABLE FOR SUPPLIER

Serial	Item Name			
		I.V. Surflo Catheter, 20G x 1"		
	Catalogue / Specification	As Attached		
(a)	(b)	(c)	(d)	(e)
	DESCRIPTION & USER REQUIREMENTS	Comply (√)	Not Comply (√)	Remark
1	Distributor Authorization Letter by the sole agency agreement from the agent			
2	Must provide manual instruction from the agent supply			
3	Clear Catalogue / Specification attached			
4	Certificate Internal Safety and Quality Standard: The supplier should specific the Product is approved by the international safety standards - TUV,FDA, CE MARK, TGA, CSA, SMG, ADA, CGSB, NF, UKAS, MS and Japanese Standard			
5	Attached SAMPLE from vendor			
	SPECIFICATION:			
6	Gauge & Length: 20G x 1"			
7	Color Code: Pink			
8	Brand: Terumo or Equivalent			
9	Product Code: SR-OX2025CA			
10	Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips			
11	Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma			
12	Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging			
13	Shelf life: Minimum 3–5 years from the date of manufacture.			
14	Quantity: 100 Boxes (box of 50 pcs)			
15	Delivery Period/Time			
16	Country Origin			
17	Quality Assurance			
Note:				
a. The product will be rejected if it does not match the specifications.				

STOR LOGISTIK PERUBATAN BOLKIAH GARISON
COMPLIANCE AGREEMENT TABLE FOR SUPPLIER

Serial	Item Name			
Catalogue / Specification		As Attached		
(a)	(b)	(c)	(d)	(e)
	DESCRIPTION & USER REQUIREMENTS	Comply (√)	Not Comply (√)	Remark
1	Distributor Authorization Letter by the sole agency agreement from the agent			
2	Must provide manual instruction from the agent supply			
3	Clear Catalogue / Specification attached			
4	Certificate Internal Safety and Quality Standard: The supplier should specific the Product is approved by the international safety standards - TUV, FDA, CE MARK, TGA, CSA, SMG, ADA, CGSB, NF, UKAS, MS and Japanese Standard			
5	Attached SAMPLE from vendor			
SPECIFICATION:				
6	Gauge & Length: 22G x 1"			
7	Color Code: Blue			
8	Brand: Terumo or Equivalent			
9	Product Code: SR-OX2225CA			
10	Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips			
11	Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma			
12	Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging			
13	Shelf life: Minimum 3–5 years from the date of manufacture.			
14	Quantity: 100 Boxes (box of 50 pcs)			
15	Delivery Period/Time			
16	Country Origin			
17	Quality Assurance			
Note:				
a. The product will be rejected if it does not match the specifications.				

STOR LOGISTIK PERUBATAN BOLKIAH GARISON
COMPLIANCE AGREEMENT TABLE FOR SUPPLIER

Serial	Item Name			
		I.V. Surflo Catheter, 24G x ¾"		
Catalogue / Specification		As Attached		
(a)	(b)	(c)	(d)	(e)
DESCRIPTION & USER REQUIREMENTS		Comply (√)	Not Comply (√)	Remark
1	Distributor Authorization Letter by the sole agency agreement from the agent			
2	Must provide manual instruction from the agent supply			
3	Clear Catalogue / Specification attached			
4	Certificate Internal Safety and Quality Standard: The supplier should specific the Product is approved by the international safety standards - TUV,FDA, CE MARK, TGA, CSA, SMG, ADA, CGSB, NF, UKAS, MS and Japanese Standard			
5	Attached SAMPLE from vendor			
SPECIFICATION:				
6	Gauge & Length: 24G x ¾"			
7	Color Code: Yellow			
8	Brand: Terumo or Equivalent			
9	Product Code: SR-OX2419CA			
10	Catheter Material: EFTE (Ethylene Tetrafluoroethylene) - known for its strength, flexibility, and resistance to split tips			
11	Needle: Ultra-sharp, stainless steel, double-bevel introducer needle with medical-grade lubricant to minimize tissue trauma			
12	Visibility: Radiopaque material ensures clear visibility under diagnostic X-ray imaging			
13	Shelf life: Minimum 3–5 years from the date of manufacture.			
14	Quantity: 100 Boxes (box of 50 pcs)			
15	Delivery Period/Time			
16	Country Origin			
17	Quality Assurance			
Note: a. The product will be rejected if it does not match the specifications.				