

KETERANGAN PEMBELIAN TEMPATAN - JKP KEMANTAN

RUJUKAN SEBUTHARGA : **MINDEF/DFA/FY2026/NAU/DEN/LP/1008**

TARIKH TUTUP SEBUTHARGA : **17 - 9 - 2026 (KHAMIS, 08:00 PAGI)**

PETI PENERIMAAN SEBUTHARGA : **BOX (3)**

ALAMAT BARU: **BANGUNAN ARKIB, ARAS BAWAH, KEMENTERIAN PERTAHANAN,
BOLKIAH GARISON**

NAMA SYARIKAT : _____

CAP SYARIKAT DAN TANDATANGAN : _____

TEL PEMBEKAL : _____ FAX PEMBEKAL : _____

E-MAIL PEMBEKAL : _____

VENDOR ID PEMBEKAL : _____

BIL	KETERANGAN BARANG	KTI	HARGA SATU	HARGA SEMUA	TEMPOH SAH HARGA (MINIMUM 180 HARI)	TEMPOH PENGHANTARAN	BUATAN NEGERI	LAIN-LAIN HAL
1	<u>PEMBELIAN PERALATAN DAN KELEGGAPAN PERGIGIAN (DENTAL EQUIPMENT)</u> ELECTROLYTE POLISHING UNIT (Sila rujuk spesifikasi di sertakan)	01 UNIT						
2	ELECTRIC WAX KNIFE (WAXLECTRIC II) (Sila rujuk spesifikasi di sertakan)	02 UNITS						
PERHATIAN: Perkara-perkara yang perlu (mandatori) di hadapkan sebelum tarikh tutup sebutharga : ☒ Katalog & Spesifikasi Teknikal mesti di sertakan untuk penilaian. ☐ Contoh Jika di dapati gagal, sebutharga tersebut tidak akan di pertimbangkan. PERINGATAN: Tidak dibenarkan menggunakan katalog Jabatanarah ini.								

Ag 27/8
JUMLAH SEMUA HARGA YANG DI TAWARKAN - ANGKA B\$ _____ DAN PERKATAAN: _____

Dental Laboratory Equipment	
Demand Reference	DSE/07/26/04
Description	Electrolyte Polishing Unit
Quantity Required	1 UNIT

Ser	Technical Data & Dimensional		Offer Model/Brand
	Item	Specification	
(a)	(b)	(c)	(d)
Essential			
1	General	An electro polisher is used in dental laboratories to refine partial denture frameworks and metal base materials. This unit employs electrochemical processes to efficiently remove surface impurities, resulting in a smoother, cleaner finish.	
1	Features and Benefits	➤ Improved to work faster and more efficient ➤ The Electrolytic Clamp & Triangular Bar are made of corrosion-proof titanium material and it makes electrolytes for a long time ➤ A built-in circuits makes it simple to operate and current can be easily supplied constantly ➤ Overload protection that isolates vibrations from the unit and main current body. ➤ Memory functions that stores (M) frequent combinations of an electrolytic power and time	
2	Specifications	Dimension (LxWxH) (203 x 242 x 133 mm) Weight 2.8kg / 6.18 lbs Power 100 W Voltages 220 – 240 V/50Hz	
Installation and Support			
11	Installation	No extra charges for the following needs; ➤ Installation. ➤ Commissioning. ➤ On-Site Demo/Training for Operators and Technical Staff.	



		➤ No limit of breakdown calls.	
12	Approved Agent Status	It is essential that the contractor/supplier submit a letter from the manufacturer stating that they: ➤ Manufacturers authorized the supplier of the equipment. ➤ Are authorized by the manufacturer to install test and commission their equipment.	
13	Equipment Whole-Life Support	It is essential that the contractor/supplier submit a letter from the manufacturer guaranteeing that spare parts will continue to be supplied for at least 10 years which is the envisaged in-service life of the equipment.	
14	Warranty	It is essential that the manufacturer provided a minimum of 2 year warranty support (parts and labor) from the date of acceptance into service.	
15	Technical Service Manual	The Supplier Shall provide during commissioning period a complete set of Technical Service manuals to assist in the conduct of routine maintenance procedures and fault finding techniques.	
Desirable			
16	Preferred Model/Brand & Country Origin of The Manufacturer	Digital Electro Polisher BesQual S730 or Equivalent.	

***Important Notice**

This specification annex form must complete and attached with the quotation tenders provided.

Company / Supplier Chop & Signature,

.....





Dental Laboratory Equipment	
Dental Services RBAF Demand Reference	DSE/07/26/04
Description	Electrolyte Polishing Unit
Quantity Required	1 Unit
Preferred Model/Brand & Country Origin of The Manufacturer	Digital Electro Polisher BesQual S730 or Equivalent.
Submitting Catalogue / Brochures	The Supplier must submit the Catalogue / Brochures completely for the offered item.



Dental Laboratory Equipment	
Demand Reference	DSE/07/26/05
Description	Electric Wax Knife (Waxelectric II)
Quantity Required	2 UNITS

Ser	Technical Data & Dimensional		Offer Model/Brand
	Item	Specification	
(a)	(b)	(c)	(d)
Essential			
1	Features and Benefits	Waxelectric II is professional device in the electric wax knife range which can be operated with two handpieces simultaneously. ➤ Temperature setting to a precise degree ensures controlled working processes. ➤ Fast heat-up rate and low heat transfer to the handle thanks to the heating element integrated in the tip. ➤ Daylight operable 3 figure LED display. ➤ Both channels can be set separately with the Waxelectric II. ➤ Color-coded for controlled management of the handpieces.	
2	Specifications	➤ Mains Voltage 220 – 240 Volts 50/60 Hz ➤ Temperature range 50 – 200 degree celcius ➤ Supply voltage 12 V DC ➤ Output Power 10 W ➤ Transformer output voltage 12 V DC	
Installation and Support			
3	Installation	No extra charges for the following needs; ➤ Installation. ➤ Commissioning. ➤ On-Site Demo/Training for Operators and Technical Staff. ➤ No limit of breakdown calls.	
4	Approved Agent Status	It is essential that the contractor/supplier submit a letter from the manufacturer stating that they: ➤ Manufacturers authorized the supplier of the equipment. ➤ Are authorized by the manufacturer to install test and commission their equipment.	
5	Equipment Whole-Life Support	It is essential that the contractor/supplier submit a letter from the manufacturer guaranteeing that spare parts will continue to be supplied for at least 10 years which is the envisaged in-service life of the equipment.	
6	Warranty	It is essential that the manufacturer provided a minimum of 2 year warranty support (parts and labor) from the date of acceptance into service.	



7	Technical Service Manual	The Supplier Shall provide during commissioning period a complete set of Technical Service manuals to assist in the conduct of routine maintenance procedures and fault finding techniques.	
Desirable			
8	Preferred Model/Brand & Country Origin of The Manufacturer	Renfert or Equivalent.	

***Important Notice**

This specification annex form must complete and attached with the quotation tenders provided.

Company / Supplier Chop & Signature,

.....





Dental Laboratory Equipment	
Dental Services RBAF Demand Reference	DSE/07/26/05
Description	Electric Wax Knife (Waxelectric II)
Quantity Required	2 Units
Preferred Model/Brand & Country Origin of The Manufacturer	Renfert or Equivalent.
Submitting Catalogue / Brochures	The Supplier must submit the Catalogue / Brochures completely for the offered item.